NIRT Wheat SRS Bias Log

Tecator Model:_____Serial # :_____ Location: _____

NOTE: Only results less than two weeks old and within a 5°F temperature range may be averaged. Results obtained after a slope or bias adjustment, SRS replacement, or NIRT repair or replacement cannot be averaged with results obtained prior to the change (also, the 5-Run correction can only be applied if five consecutive, valid runs are all positive or all negative). Every effort should be made to maintain the bias within the 0.05% Protein tolerance.

Date mm/dd/yy	Temp	**Protein Class:**						Oper Init	**Comments**
		RH	Today's Bias	Multi-Run Averages			S* B* R*		
				2-Run	3-Run	5-Run			
			± 0.10	± 0.07	± 0.05	± 0.03			

* (S) slope adjustment, (B) bias adjustment, or (R) SRS replaced. If the NIRT is repaired or replaced, start a new log.