NIRT Soybean SRS Bias Log

Tecator Model:_____Serial # : _____ Location:_____

NOTE: Only results less than two weeks old and within a 5°F temperature range may be averaged. Results obtained after a slope or bias adjustment, SRS replacement, or NIRT repair or replacement cannot be averaged with results obtained prior to the change (also, the 5-Run correction can only be applied if five consecutive, valid runs are all positive or all negative). Every effort should be made to maintain the bias within the 0.08% Protein and 0.05% Oil tolerances.

Date mm/dd/yy	Temp	**Protein**						**Oil**					
		RH	Today's Bias	Multi-Run Averages			S* B* R*	Today's Bias	Multi-Run Averages			S* B* R*	Oper Init
				2-Run	3-Run	5-Run			2-Run	3-Run	5-Run		
			±0.17	±0.12	±0.10	±0.08		±0.12	±0.09	±0.07	±0.05		

*(S) slope adjustment, (B) bias adjustment, or (R) SRS replaced. If the NIRT is repaired or replaced, start a new log.